DISPOSABLE FACE MASK
3-Ply Earloop Face Mask Classic Style | BLUE
CE ISO
DISPOSABLE FACE MASK
50 PER BOX
3-Ply Earloop Face Mask
Classic Style | BLUE
DO NOT REUSE
DISPOSABLE FACE MASK
50 PER BOX
3-Ply Earloop Face Mask



Certificate

No. ICR Polska/M8403020

Name and address of certificate owner:	Jiangxi Kangmeijie Hygiene Materials Co.,Ltd No.888 Lianfu Avenue, Changleng Industry Zone, Xinjian District, Nanchang City, Jiangxi Province, PRC
Name and address of manufacturer:	Jiangxi Kangmeijie Hygiene Materials Co.,Ltd No.888 Lianfu Avenue, Changleng Industry Zone, Xinjian District, Nanchang City, Jiangxi Province, PRC
Product name:	Disposable Face Mask
Product types:	FM01,FM02,FM03,FM04,FM05,FM06,FM07,FM08,FM09

This certificate confirms that the product meets the requirements of the following standards and within limits of its standards gives presumption of conformity with essential requirements of Regulation 2016/425

EN 149:2001+A1:2009

The certification process has been carried out in accordance with the program PC-P-07-07.

Evaluation has been carried out in accordance with test reports made by EUROPEAN QUALITY TEST CO., LTD.

No. of test reports: KML-2020-0314

Certificate issue date: 18.03.2020
Expiration date: 17.03.2025

The mutual obligations and rights of the certification are regulated by the contract No. ICR Polska/2020-8458.

This certificate applies to products having the same attributes (parameters), intended use, that have been evaluated and meet the requirements of the aforementioned standards.

Director: Rafał Kalinowski

Warsaw, 18.03.2020

ICR Polska Co. Ltd.
ul. Plac Przymierza 6, 03-944 Warszawa
www.icrpolska.com, e-mail: icrpolska@icrqa.com



intertek
Total Quality. Assured.

CERTIFICATE OF REGISTRATION

This is to certify that the management system of:

[redacted]
MEDICAL CO., LTD.

Main Site: [redacted] Jiangsu Province, China

has been registered by Intertek as conforming to the requirements of:

ISO 13485:2016

The management system is applicable to:

Manufacture and distribution of disposable plastic medical equipment covers, non-woven drapes, Urine Collection bag, wound bed preparation tool, disposable plastic medical products such as fluid management system connectors, valves, spike, roller clamp, hand brush, inhalers, blood smear applicator, type safe segment piercing device, blood transfer straw, urine transfer straw, urine tube and cap, negative suction system.

Certificate Number:
0085778

Initial Certification Date:
15 July 2011

Certificate Issue Date:
04 January 2019

Certificate Expiry Date:
14 July 2020

Calin Moldovean
President, Business Assurance

Intertek Testing Services NA Ltd.,
1829, 32nd avenue, Lachine, QC, H8T 3J1,
Canada



PONY 谱尼测试
Pony Testing International Group

Test Report

No. [redacted]

Sample Description: Disposable face mask for medical use

Applicant: [redacted]

Pony Testing International Group
www.ponytest.com



PONY 谱尼测试
Pony Testing International Group

Test Results

No. BOYIRGIRT81487QA Page 1 of 2

Sample Description	Disposable face mask for medical use	Sample Specification	——
Applicant	[redacted]	Trade Mark	[redacted]
Received Date	2020-03-17	Manufacturing Date or Lot No.	2020.3.9 20200309
Test Date	2020-03-17~2020-03-27	Sample Grade	——
Sample Quantity	66Pcs	Test Type	Commissioning Test
Sample Status	Solid	Test Environment	To meet the requirements
Test Items	See the next page		
Reference Methods	See the next page		
Main Instruments	Bacterial filtration efficiency detector、Constant temperature incubator etc.		
Note	Limit on：EN 14683：2019 Sample source：self-made		
	Edited by	Lu Wenghu	
	Checked by	Shao	
	Approved by	Daiqing	
	Issued Date	2020-03-27	

Hotline 400-819-5688
www.ponytest.com



PONY 谱尼测试
Pony Testing International Group

Test Results

No. BOYIRGIRT81487QA Page 2 of 2

Sample Description and Number	Test Items	Limit			Test Results		Reference Methods
		Type I	Type II	Type IIR			
RT81487QA Disposable face mask for medical use	Bacterial filtration efficiency(BFE)，%	≥95	≥98	≥98	No.1	98.4	EN 14683：2019
					No.2	98.8	
					No.3	99.0	
					No.4	98.5	
					No.5	99.1	
	Differential pressure，Pa/cm²	<40	<40	<60	31		EN 14683：2019
	Splash resistance pressure，kPa	/	/	≥16.0	>16.0		EN 14683：2019
	Microbial cleanliness，cfu/g	≤30	≤30	≤30	4		EN 14683：2019

Sample Number and Photo：

RT81487QA

Pony authenticate the photo on original report only
——End of Report——

Hotline 400-819-5688
www.ponytest.com